

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 22 2009

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130114 ✓

Name Kevin Stevenson

ID# 2414 Date 09-18-09 ✓

Agency UAA Police Dept

Phone # 907-786-1120

Instrument Location UAA 3211 Providence Dr. ESH 114 ✓

Alco S/N 75470 ✓

Target Value .080 ✓

High Pressure 800

Alco Test Value Average .080
1st Alco

.080
2nd Alco ✓

Signature [Signature] ✓

(OVER)

OK
9/22/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz

Scientific Director


State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of Oct, 2009.

Carolyn M. Noland
Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

SEPTEMBER 18, 2009

OPERATOR'S NAME:
STEVENSON KEVIN M
OPERATOR'S NUMBER: 2414
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
VERIFICATION
O.L. #: 205432
DEPT/AGENCY: DMC7
CASE/REPORT: 090907
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: 75470

— BREATH ANALYSIS —

.000 ADJUSTED FOR 29.55 in
ALCO TARGET .079 16:57
BLANK TEST .000 16:58
INTERNAL STANDARD VERIFIED 16:58
ALCO TO 29.55 in .000 16:58
BLANK TEST .000 16:59
SUBJECT SAMPLE .000 17:00
BLANK TEST .000 17:01
ALCO TO 29.57 in .000 17:01
BLANK TEST .000 17:02

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

SEPTEMBER 18, 2009
TIME 16:54

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS:
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.55 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTION: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~@